
Midmark M11 - v1.0.5
Total Cycles: 1095

Name: RITTER2
Sterilizer ID: V2313228
05/11/2022 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 183.9 F 4.0
4:00 192.4 F 9.0
6:00 222.2 F 7.5
8:00 233.0 F 10.8
10:00 234.7 F 8.4
12:00 239.7 F 10.4
14:00 247.4 F 13.9
16:00 259.7 F 20.7
18:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.3 F 29.1
2:00 272.6 F 29.8
3:00 272.1 F 29.1
4:00 271.8 F 29.7

Min 270.3 F 28.3
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 212.7 F 0.0
10:00 218.5 F 0.0
15:00 230.6 F 0.0
20:00 227.8 F 0.0
25:00 225.4 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

05/11/2022 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____