
Midmark M11 - v1.0.0
Total Cycles: 5139

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 152.3 F 0.1
4:00 156.2 F 0.2
6:00 183.3 F 0.7
8:00 203.6 F 3.6
10:00 216.4 F 7.5
12:00 230.3 F 12.0
14:00 231.7 F 8.2
16:00 243.2 F 12.2
18:00 243.1 F 12.2
20:00 241.4 F 11.1
22:00 254.8 F 17.7
24:00 265.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 271.7 F 28.5
2:00 272.5 F 28.9
3:00 272.1 F 28.5
4:00 271.3 F 28.2

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 210.3 F 0.0
10:00 218.9 F 0.0
15:00 237.1 F 0.0
20:00 236.6 F 0.0
25:00 229.3 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

11/28/2024 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____