
Midmark M11 - v1.0.0
Total Cycles: 3480

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 191.9 F 3.6
4:00 199.9 F 8.2
6:00 219.1 F 11.3
8:00 228.2 F 8.7
10:00 236.8 F 10.8
12:00 241.1 F 11.6
14:00 239.1 F 10.0
16:00 252.7 F 16.7
18:00 264.3 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.6 F 28.3
3:00 272.3 F 29.1
4:00 272.3 F 28.7

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 216.7 F 0.0
10:00 231.8 F 0.0
15:00 228.3 F 0.0
20:00 223.6 F 0.0
25:00 223.0 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

07/04/2023 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____