
Midmark M11 - v1.0.0
Total Cycles: 4636

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2024 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.8 F 0.0
2:00 137.1 F 1.7
4:00 152.5 F 4.6
6:00 174.1 F 7.6
8:00 204.2 F 11.4
10:00 216.6 F 3.5
12:00 220.2 F 7.3
14:00 230.8 F 11.2
16:00 229.1 F 6.8
18:00 242.0 F 12.0
20:00 239.0 F 10.0
22:00 251.0 F 15.8
24:00 261.4 F 21.6
26:00 270.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 29.2
2:00 271.5 F 28.5
3:00 272.8 F 29.0
4:00 271.7 F 28.9

Min 271.0 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 228.6 F 0.0
10:00 243.2 F 0.0
15:00 201.7 F 0.0
20:00 181.8 F 0.0
25:00 176.9 F 0.0
30:00 174.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:26
CYCLE COMPLETE

06/13/2024 12:30 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____