
Midmark M11 - v1.0.0
Total Cycles: 3627

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 189.3 F 4.3
4:00 212.1 F 10.1
6:00 222.7 F 5.6
8:00 236.0 F 11.4
10:00 235.5 F 8.9
12:00 239.1 F 9.7
14:00 248.3 F 14.5
16:00 260.7 F 21.3
18:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.7 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 215.5 F 0.0
10:00 231.7 F 0.0
15:00 229.7 F 0.0
20:00 223.5 F 0.0
25:00 220.7 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

08/09/2023 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____