
Midmark M11 - v1.0.5
Total Cycles: 3744

Name: RITTER2
Sterilizer ID: V2313228
06/28/2024 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.8 F 0.0
2:00 141.7 F 1.5
4:00 149.8 F 4.4
6:00 176.7 F 8.1
8:00 197.7 F 12.3
10:00 211.4 F 5.8
12:00 220.9 F 10.4
14:00 229.2 F 7.9
16:00 237.2 F 12.1
18:00 238.9 F 10.5
20:00 251.7 F 16.6
22:00 262.5 F 23.0
24:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 271.8 F 29.5
2:00 271.8 F 29.4
3:00 272.1 F 29.6
4:00 271.9 F 29.8

Min 271.4 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 222.0 F 0.0
10:00 224.8 F 0.0
15:00 198.7 F 0.0
20:00 192.7 F 0.0
25:00 192.5 F 0.0
30:00 191.1 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:03
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:40
CYCLE COMPLETE

06/28/2024 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____