
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007950
8:49 09/10/2024

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	2:45
PRESSURIZING	4:56
STERILIZING	5:43
136.0 C 220kPa	5:43
136.4 C 223kPa	6:13
136.2 C 221kPa	6:43
136.4 C 223kPa	7:13
136.2 C 221kPa	7:43
136.3 C 222kPa	8:13
136.4 C 222kPa	8:43
136.3 C 222kPa	9:13
136.4 C 224kPa	9:43
136.2 C 220kPa	10:13
136.4 C 223kPa	10:43
136.4 C 221kPa	11:13
136.4 C 222kPa	11:43
136.4 C 220kPa	12:13
136.4 C 218kPa	12:43
136.2 C 218kPa	13:13
136.4 C 223kPa	13:43
136.4 C 222kPa	14:13
136.1 C 219kPa	14:43
136.3 C 221kPa	15:13
136.3 C 223kPa	15:43
136.3 C 223kPa	15:43
VENTING	15:43
AIR DRYING	16:34
CYCLE COMPLETE	17:41

Digital Signature #

14F49EFB149F1E95
