
Midmark M11 - v1.0.5
Total Cycles: 4271

Name: RITTER2
Sterilizer ID: V2313228
01/16/2025 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.0 F 0.0
2:00 94.9 F 0.0
4:00 112.9 F 0.0
6:00 145.2 F 0.3
8:00 168.4 F 1.4
10:00 191.3 F 4.6
12:00 208.2 F 8.1
14:00 218.9 F 12.0
16:00 225.9 F 7.4
18:00 236.7 F 12.0
20:00 234.9 F 9.4
22:00 239.7 F 10.4
24:00 245.6 F 13.6
26:00 256.8 F 19.6
28:00 266.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.5 F 29.7
2:00 271.4 F 29.0
3:00 272.4 F 29.8
4:00 271.9 F 29.2

Min 271.1 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 235.7 F 0.0
10:00 209.4 F 0.0
15:00 188.5 F 0.0
20:00 181.2 F 0.0
25:00 179.7 F 0.0
30:00 179.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 29:36
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:08:25
CYCLE COMPLETE

01/16/2025 04:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____