

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002396  
12:03 18/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:31  
PRESSURIZING 1:37  
STERILIZING 2:45  
136.4 C 222kPa 2:45  
136.4 C 225kPa 3:15  
136.5 C 223kPa 3:45  
136.2 C 221kPa 4:15  
136.2 C 223kPa 4:45  
136.2 C 223kPa 5:15  
136.2 C 222kPa 5:45  
136.3 C 225kPa 6:15  
136.3 C 225kPa 6:15  
VENTING 6:15  
AIR DRYING 7:05  
CYCLE COMPLETE 17:02  
-----

Digital Signature #  
356BF82F9128042B  
-----