
Midmark M11 - v1.0.0
Total Cycles: 1566

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2022 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 176.1 F 3.7
4:00 188.0 F 8.4
6:00 214.2 F 9.1
8:00 227.7 F 8.4
10:00 239.1 F 11.8
12:00 239.8 F 11.0
14:00 237.8 F 9.2
16:00 251.9 F 16.0
18:00 263.5 F 22.9
20:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 28.9
2:00 271.9 F 28.3
3:00 271.5 F 28.3
4:00 272.7 F 29.0

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 233.9 F 0.0
10:00 223.1 F 0.0
15:00 223.9 F 0.0
20:00 225.0 F 0.0
25:00 227.4 F 0.0
30:00 232.5 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

01/26/2022 12:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____