
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002328
8:24 29/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 1:55
STOP BUTTON PRESSED
NOT STERILE
END TIME 1:55

Digital Signature #
32036315116A2188
