

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3297

Name: RITTER 1  
Sterilizer ID: V2178198  
05/29/2023 12:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
2:00 189.8 F 0.6  
4:00 208.8 F 6.0  
6:00 221.6 F 11.4  
8:00 229.2 F 7.4  
10:00 240.9 F 13.3  
12:00 241.7 F 11.5  
14:00 240.3 F 10.4  
16:00 254.4 F 17.5  
18:00 266.3 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.4 F 29.0  
2:00 272.3 F 28.6  
3:00 271.5 F 28.0  
4:00 271.4 F 28.4  
  
Min 271.2 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 217.9 F 0.0  
10:00 220.3 F 0.0  
15:00 218.3 F 0.0  
20:00 221.2 F 0.0  
25:00 226.9 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 1:49  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

05/29/2023 01:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_