
Midmark M11 - v1.0.0
Total Cycles: 621

Name: MIDMARK
Sterilizer ID: V2178198
01/21/2021 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 190.1 F 3.4
4:00 203.3 F 8.8
6:00 227.5 F 8.6
8:00 229.3 F 10.1
10:00 233.7 F 8.4
12:00 243.5 F 12.1
14:00 245.0 F 13.0
16:00 257.6 F 19.6
18:00 268.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.7 F 29.2
2:00 271.2 F 28.6
3:00 272.3 F 29.3
4:00 271.5 F 28.6

Min 271.1 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 241.2 F 0.0
10:00 221.5 F 0.0
15:00 224.8 F 0.0
20:00 228.0 F 0.0
25:00 232.4 F 0.0
Dry Cycle Aborted

FILLING: 2:25
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:20
DRYING: 27:48
TOTAL CYCLE: 00:56:42

01/21/2021 05:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____