
Midmark M11 - v1.0.0
Total Cycles: 848

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 06:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 73.0 F 0.0
4:00 100.5 F 0.1
6:00 123.8 F 0.1
8:00 144.4 F 0.2
10:00 183.4 F 0.3
12:00 205.6 F 3.2
14:00 214.5 F 6.1
16:00 224.1 F 9.2
18:00 233.5 F 13.0
20:00 233.6 F 8.5
22:00 244.2 F 13.1
24:00 242.2 F 11.3
26:00 238.8 F 9.5
28:00 252.9 F 16.3
30:00 264.4 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 271.4 F 27.9
2:00 271.5 F 27.9
3:00 271.4 F 27.9
4:00 271.4 F 27.8

Min 271.3 F 27.7
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 214.1 F 0.0
10:00 233.3 F 0.0
15:00 221.3 F 0.0
20:00 210.3 F 0.0
25:00 205.7 F 0.0
30:00 205.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:51
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

05/31/2021 07:44 AM
ACCEPTED: X
REJECTED:
Operator: U01