
Midmark M11 - v1.0.5
Total Cycles: 3702

Name: RITTER2
Sterilizer ID: V2313228
06/14/2024 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.5 F 0.0
2:00 114.0 F 0.6
4:00 130.8 F 2.7
6:00 156.9 F 5.6
8:00 178.0 F 8.8
10:00 192.9 F 12.3
12:00 215.8 F 5.0
14:00 229.1 F 8.7
16:00 229.7 F 12.9
18:00 233.2 F 9.3
20:00 240.0 F 11.1
22:00 244.0 F 12.7
24:00 254.9 F 18.3
26:00 264.9 F 24.4
28:00 270.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.7 F 29.8
2:00 272.2 F 30.1
3:00 272.0 F 29.6
4:00 272.6 F 30.2

Min 271.5 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 232.1 F 0.0
10:00 187.7 F 0.0
15:00 182.1 F 0.0
20:00 176.1 F 0.0
25:00 173.7 F 0.0
30:00 172.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:52
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

06/14/2024 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____