
Midmark M11 - v1.0.0
Total Cycles: 5283

Name: RITTER 1
Sterilizer ID: V2178198
01/27/2025 07:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.2 F 0.0
2:00 114.6 F 0.5
4:00 133.9 F 3.0
6:00 161.0 F 5.9
8:00 183.2 F 9.2
10:00 199.8 F 13.0
12:00 218.5 F 6.0
14:00 228.2 F 10.4
16:00 229.6 F 7.7
18:00 238.3 F 10.9
20:00 234.1 F 8.1
22:00 247.5 F 14.2
24:00 258.4 F 20.3
26:00 268.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.4
2:00 271.4 F 29.1
3:00 272.6 F 29.5
4:00 271.4 F 28.9

Min 271.3 F 28.9
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 211.2 F 0.0
10:00 226.5 F 0.0
15:00 230.3 F 0.0
20:00 223.8 F 0.0
25:00 227.3 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:60
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:06:50
CYCLE COMPLETE

01/27/2025 08:32 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____