

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001362  
12:23 24/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:25  
PRESSURIZING 4:26  
STERILIZING 5:32  
136.4 C 224kPa 5:32  
136.4 C 225kPa 6:02  
136.3 C 224kPa 6:32  
136.2 C 222kPa 7:02  
136.3 C 224kPa 7:32  
136.3 C 224kPa 8:02  
136.2 C 223kPa 8:32  
136.3 C 224kPa 9:02  
136.3 C 224kPa 9:02  
VENTING 9:02  
AIR DRYING 9:57  
CYCLE COMPLETE 24:32  
-----

Digital Signature #  
038638F0E9C73409  
-----