
Midmark M11 - v1.0.5
Total Cycles: 4298

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.9 F 0.0
2:00 115.2 F 0.1
4:00 115.7 F 1.9
6:00 133.1 F 4.7
8:00 157.0 F 8.0
10:00 182.7 F 11.7
12:00 200.7 F 3.9
14:00 211.9 F 7.8
16:00 225.6 F 11.9
18:00 230.4 F 8.5
20:00 240.2 F 13.5
22:00 241.1 F 12.0
24:00 252.6 F 17.7
26:00 262.8 F 23.7
28:00 272.6 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.5
1:00 272.3 F 30.6
2:00 271.9 F 30.1
3:00 272.6 F 30.5
4:00 272.1 F 30.7

Min 270.8 F 29.4
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 209.6 F 0.0
10:00 209.2 F 0.0
15:00 194.4 F 0.0
20:00 183.7 F 0.0
25:00 190.1 F 0.0
30:00 193.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:28
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:07:19
CYCLE COMPLETE

01/27/2025 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____