
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001039
12:19 22/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:58
136.5 C 225kPa 4:58
136.4 C 225kPa 5:28
136.4 C 224kPa 5:58
136.2 C 223kPa 6:28
136.1 C 222kPa 6:58
136.4 C 225kPa 7:28
136.3 C 223kPa 7:58
136.4 C 224kPa 8:28
136.4 C 224kPa 8:28
VENTING 8:28
AIR DRYING 9:23
CYCLE COMPLETE 9:25

Digital Signature #
09E8640BAFD5EF62
