

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001754  
16:06 24/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:40  
STERILIZING 3:37  
136.0 C 221kPa 3:37  
136.4 C 225kPa 4:07  
136.4 C 223kPa 4:37  
136.4 C 225kPa 5:07  
136.4 C 225kPa 5:37  
136.3 C 224kPa 6:07  
136.1 C 221kPa 6:37  
136.4 C 223kPa 7:07  
136.4 C 223kPa 7:07  
VENTING 7:07  
AIR DRYING 8:00  
CYCLE COMPLETE 11:14

-----  
Digital Signature #  
6D419E968CE6C67C  
-----