
Midmark M11 - v1.0.0
Total Cycles: 2320

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 192.1 F 4.0
4:00 207.2 F 9.1
6:00 224.8 F 7.3
8:00 233.3 F 11.1
10:00 234.1 F 8.3
12:00 240.1 F 10.0
14:00 246.6 F 13.6
16:00 259.5 F 20.7
18:00 270.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.0 F 28.9
2:00 272.1 F 29.2
3:00 272.5 F 29.4
4:00 272.4 F 29.6

Min 271.6 F 28.9
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 239.5 F 0.0
10:00 227.3 F 0.0
15:00 229.2 F 0.0
20:00 230.0 F 0.0
25:00 227.4 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 2:52
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

09/13/2022 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____