
Midmark M11 - v1.0.0
Total Cycles: 4580

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2024 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.6 F 0.0
2:00 132.9 F 1.7
4:00 153.5 F 4.5
6:00 173.9 F 7.6
8:00 190.5 F 11.3
10:00 214.1 F 3.8
12:00 217.3 F 7.7
14:00 228.4 F 11.7
16:00 229.5 F 7.4
18:00 240.7 F 12.4
20:00 238.9 F 10.2
22:00 250.6 F 15.9
24:00 260.9 F 21.7
26:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.4
2:00 272.5 F 29.1
3:00 271.5 F 28.6
4:00 271.3 F 28.7

Min 271.2 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 204.9 F 0.0
10:00 207.1 F 0.0
15:00 189.8 F 0.0
20:00 199.5 F 0.0
25:00 209.9 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:49
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:31
CYCLE COMPLETE

05/24/2024 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____