
Midmark M11 - v1.0.5
Total Cycles: 1553

Name: RITTER2
Sterilizer ID: V2313228
10/03/2022 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.5 F 0.0
2:00 157.0 F 0.0
4:00 186.2 F 0.4
6:00 208.4 F 4.0
8:00 225.6 F 8.8
10:00 236.8 F 11.1
12:00 237.4 F 11.2
14:00 237.3 F 9.5
16:00 239.4 F 9.2
18:00 250.9 F 15.9
20:00 263.1 F 23.0
22:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.4
2:00 271.7 F 28.6
3:00 272.3 F 29.5
4:00 272.3 F 29.1

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 229.0 F 0.0
10:00 247.3 F 0.0
15:00 228.3 F 0.0
20:00 224.0 F 0.0
25:00 220.7 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

10/03/2022 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____