
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000596
14:39 20/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:42
STERILIZING 3:47
136.4 C 231kPa 3:47
136.2 C 232kPa 4:17
136.3 C 231kPa 4:47
136.2 C 232kPa 5:17
136.4 C 233kPa 5:47
136.3 C 232kPa 6:17
136.2 C 232kPa 6:47
136.4 C 232kPa 7:17
136.4 C 232kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CASSETTE REMOVED
CYCLE COMPLETE 10:59

Digital Signature #
AEFF9765014FC004
