
Midmark M11 - v1.0.5
Total Cycles: 2077

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 171.7 F 2.3
4:00 192.7 F 6.5
6:00 210.1 F 12.0
8:00 224.0 F 7.6
10:00 238.5 F 13.5
12:00 240.6 F 11.8
14:00 240.8 F 11.1
16:00 254.9 F 18.3
18:00 267.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.5
2:00 271.4 F 28.8
3:00 272.3 F 29.7
4:00 272.1 F 29.2

Min 271.1 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 230.7 F 0.0
10:00 212.8 F 0.0
15:00 226.8 F 0.0
20:00 229.2 F 0.0
25:00 224.9 F 0.0
30:00 224.6 F 0.0

Drying Complete

FILLING: 1:14
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

02/22/2023 02:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____