
Midmark M11 - v1.0.0
Total Cycles: 5215

Name: RITTER 1
Sterilizer ID: V2178198
01/02/2025 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.5 F 0.0
4:00 83.9 F 0.0
6:00 114.8 F 0.0
8:00 147.6 F 0.1
10:00 173.2 F 0.2
12:00 192.4 F 0.3
14:00 204.7 F 0.7
16:00 216.8 F 3.3
18:00 223.1 F 6.0
20:00 231.3 F 9.1
22:00 240.5 F 12.9
24:00 238.2 F 9.9
26:00 238.3 F 9.3
28:00 245.5 F 12.3
30:00 245.7 F 12.9
32:00 257.1 F 19.0
34:00 266.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.5 F 28.5
2:00 272.1 F 28.5
3:00 272.1 F 28.2
4:00 272.6 F 28.8

Min 271.6 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 154.6 F 0.0
5:00 205.3 F 0.0
10:00 208.0 F 0.0
15:00 201.2 F 0.0
20:00 201.9 F 0.0
25:00 206.4 F 0.0
30:00 209.1 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 35:38
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:14:06
CYCLE COMPLETE

01/02/2025 10:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____