
Midmark M11 - v1.0.0
Total Cycles: 4430

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2024 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.6 F 0.0
2:00 138.4 F 1.7
4:00 154.3 F 4.5
6:00 178.7 F 8.0
8:00 197.1 F 11.9
10:00 216.5 F 5.3
12:00 225.2 F 9.5
14:00 235.3 F 10.9
16:00 238.4 F 10.3
18:00 235.9 F 8.6
20:00 247.1 F 13.8
22:00 257.9 F 19.7
24:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.5 F 29.0
2:00 272.3 F 29.2
3:00 272.1 F 29.3
4:00 271.2 F 29.0

Min 271.0 F 28.7
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 202.2 F 0.0
10:00 181.4 F 0.0
15:00 178.0 F 0.0
20:00 176.2 F 0.0
25:00 176.0 F 0.0
30:00 175.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:58
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:42
CYCLE COMPLETE

04/05/2024 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____