
Midmark M11 - v1.0.5
Total Cycles: 4280

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 156.9 F 2.5
4:00 174.6 F 6.5
6:00 194.2 F 11.1
8:00 208.2 F 4.6
10:00 223.2 F 9.3
12:00 232.9 F 10.7
14:00 238.9 F 11.7
16:00 237.9 F 10.3
18:00 250.9 F 16.6
20:00 262.2 F 23.0
22:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.8 F 29.8
2:00 271.4 F 29.2
3:00 272.5 F 30.0
4:00 271.8 F 29.4

Min 271.3 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 201.4 F 0.0
10:00 193.6 F 0.0
15:00 194.9 F 0.0
20:00 199.6 F 0.0
25:00 203.2 F 0.0
30:00 205.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

01/22/2025 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____