
Midmark M11 - v1.0.0
Total Cycles: 2977

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 172.0 F 1.5
4:00 178.5 F 5.2
6:00 206.1 F 10.0
8:00 218.6 F 6.3
10:00 233.6 F 9.5
12:00 235.9 F 9.4
14:00 242.8 F 12.2
16:00 240.7 F 10.5
18:00 254.1 F 17.3
20:00 265.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 271.9 F 28.2
2:00 272.5 F 28.7
3:00 271.5 F 28.4
4:00 271.8 F 28.3

Min 271.3 F 28.0
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 233.9 F 0.0
10:00 220.2 F 0.0
15:00 215.0 F 0.0
20:00 213.4 F 0.0
25:00 213.5 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

03/08/2023 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____