
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008259
15:47 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:16
STERILIZING 4:20
136.0 C 220kPa 4:20
136.4 C 223kPa 4:50
136.1 C 223kPa 5:20
136.0 C 221kPa 5:50
136.4 C 223kPa 6:20
136.4 C 225kPa 6:50
136.4 C 225kPa 7:20
136.4 C 224kPa 7:50
136.4 C 224kPa 7:50
VENTING 7:50
AIR DRYING 8:49
CYCLE COMPLETE 9:26

Digital Signature #
5BF78033E58D68DC
