
Midmark M11 - v1.0.0
Total Cycles: 5092

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 166.0 F 0.4
4:00 183.4 F 0.9
6:00 203.4 F 4.2
8:00 220.0 F 8.3
10:00 229.1 F 13.1
12:00 234.0 F 9.0
14:00 239.5 F 10.4
16:00 244.3 F 12.6
18:00 243.0 F 11.6
20:00 255.7 F 18.1
22:00 266.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.6
2:00 271.3 F 27.9
3:00 272.4 F 28.8
4:00 271.6 F 28.2

Min 271.2 F 27.9
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 221.1 F 0.0
10:00 223.1 F 0.0
15:00 228.5 F 0.0
20:00 230.6 F 0.0
25:00 230.0 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

11/14/2024 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____