
Midmark M11 - v1.0.0
Total Cycles: 1284

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2021 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.1 F 0.0
2:00 173.7 F 2.8
4:00 186.8 F 6.8
6:00 210.5 F 12.5
8:00 221.8 F 6.9
10:00 235.1 F 12.0
12:00 235.0 F 9.2
14:00 238.5 F 9.7
16:00 246.9 F 13.8
18:00 259.1 F 20.5
20:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 29.0
2:00 271.7 F 28.6
3:00 272.7 F 29.1
4:00 271.6 F 28.4

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 195.0 F 0.0
10:00 204.8 F 0.0
15:00 206.4 F 0.0
20:00 211.2 F 0.0
25:00 206.3 F 0.0
30:00 203.1 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

10/21/2021 10:18 AM
ACCEPTED: X
REJECTED:
Operator: U02