
Midmark M11 - v1.0.5
Total Cycles: 1620

Name: RITTER2
Sterilizer ID: V2313228
10/19/2022 07:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 79.8 F 0.0
4:00 110.3 F 0.0
6:00 117.9 F 0.0
8:00 147.1 F 0.1
10:00 179.3 F 0.2
12:00 201.8 F 3.2
14:00 212.8 F 6.4
16:00 223.5 F 9.9
18:00 232.6 F 11.5
20:00 235.2 F 10.0
22:00 237.1 F 9.6
24:00 244.9 F 13.2
26:00 243.2 F 12.1
28:00 255.1 F 18.4
30:00 265.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.4
2:00 271.3 F 28.6
3:00 272.4 F 29.6
4:00 271.8 F 29.0

Min 271.2 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 202.9 F 0.0
10:00 208.6 F 0.0
15:00 208.2 F 0.0
20:00 216.3 F 0.0
25:00 216.3 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:54
CYCLE COMPLETE

10/19/2022 08:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____