
Midmark M11 - v1.0.5
Total Cycles: 282

Name: RITTER2
Sterilizer ID: V2313228
08/18/2021 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.4 F 0.0
2:00 156.1 F 1.4
4:00 167.6 F 4.3
6:00 189.4 F 7.9
8:00 203.7 F 11.9
10:00 214.9 F 6.0
12:00 223.6 F 10.3
14:00 231.9 F 8.8
16:00 237.7 F 11.9
18:00 238.7 F 10.3
20:00 250.2 F 16.0
22:00 260.3 F 21.9
24:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.9
1:00 271.7 F 29.9
2:00 272.4 F 30.6
3:00 271.5 F 30.3
4:00 272.1 F 30.4

Min 270.9 F 29.8
Max 273.2 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 236.2 F 0.0
10:00 192.6 F 0.0
15:00 185.6 F 0.0
20:00 180.9 F 0.0
25:00 179.4 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 25:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:09
CYCLE COMPLETE

08/18/2021 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____