
Midmark M11 - v1.0.5
Total Cycles: 4029

Name: RITTER2
Sterilizer ID: V2313228
10/15/2024 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.8 F 0.0
2:00 88.9 F 0.0
4:00 113.0 F 0.0
6:00 147.4 F 0.4
8:00 173.8 F 2.9
10:00 189.5 F 6.1
12:00 202.9 F 9.7
14:00 216.0 F 13.5
16:00 222.2 F 7.2
18:00 233.4 F 11.6
20:00 233.4 F 8.9
22:00 240.7 F 11.1
24:00 245.7 F 13.7
26:00 257.3 F 19.9
28:00 267.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 273.0 F 29.7
2:00 271.8 F 29.0
3:00 273.0 F 29.9
4:00 272.3 F 29.3

Min 271.7 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 198.2 F 0.0
10:00 201.8 F 0.0
15:00 202.6 F 0.0
20:00 201.6 F 0.0
25:00 201.9 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

10/15/2024 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____