
Midmark M11 - v1.0.0
Total Cycles: 2837

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 04:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.6 F 0.0
2:00 200.9 F 4.5
4:00 204.5 F 9.6
6:00 219.8 F 6.6
8:00 230.1 F 9.9
10:00 236.6 F 8.1
12:00 244.4 F 13.2
14:00 243.6 F 12.0
16:00 256.7 F 19.0
18:00 268.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 28.9
2:00 271.3 F 28.4
3:00 272.4 F 29.0
4:00 271.5 F 28.3

Min 271.2 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 216.8 F 0.0
10:00 234.6 F 0.0
15:00 231.2 F 0.0
20:00 228.3 F 0.0
25:00 228.5 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

01/31/2023 05:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____