
Midmark M11 - v1.0.0
Total Cycles: 1424

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2021 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.1 F 0.0
2:00 192.8 F 2.6
4:00 209.0 F 7.9
6:00 224.7 F 11.7
8:00 234.0 F 9.6
10:00 235.8 F 8.9
12:00 245.7 F 13.4
14:00 244.0 F 12.4
16:00 257.3 F 19.3
18:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.9 F 28.5
2:00 272.4 F 28.9
3:00 271.7 F 28.7
4:00 271.6 F 28.5

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 213.2 F 0.0
10:00 185.2 F 0.0
15:00 178.4 F 0.0
20:00 187.6 F 0.0
25:00 169.2 F 0.0
30:00 167.9 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

12/06/2021 12:47 PM
ACCEPTED: X
REJECTED:
Operator: U02