

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2808

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2023 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
2:00 190.7 F 3.3  
4:00 199.2 F 8.3  
6:00 219.0 F 9.8  
8:00 229.5 F 9.2  
10:00 235.2 F 9.4  
12:00 243.1 F 12.5  
14:00 241.6 F 11.2  
16:00 255.0 F 18.0  
18:00 266.5 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.0 F 29.1  
2:00 271.9 F 28.7  
3:00 271.4 F 28.6  
4:00 272.3 F 29.1  
  
Min 271.1 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 227.4 F 0.0  
10:00 231.1 F 0.0  
15:00 229.4 F 0.0  
20:00 227.8 F 0.0  
25:00 227.5 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

01/23/2023 05:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_