
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007968
12:18 11/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:34
STERILIZING 3:30
136.0 C 220kPa 3:30
136.2 C 223kPa 4:00
136.1 C 222kPa 4:30
136.3 C 222kPa 5:00
135.9 C 219kPa 5:30
136.3 C 221kPa 6:00
136.2 C 223kPa 6:30
136.4 C 224kPa 7:00
136.4 C 224kPa 7:00
VENTING 7:00
AIR DRYING 7:47
CYCLE COMPLETE 37:48

Digital Signature #
70E0A013F018C6BD
