

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000162  
13:10 15/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:28  
PRESSURIZING 4:28  
STERILIZING 5:44  
136.0 C 220kPa 5:44  
136.4 C 224kPa 6:15  
136.3 C 225kPa 6:45  
136.3 C 226kPa 7:15  
136.1 C 223kPa 7:45  
136.3 C 224kPa 8:15  
136.3 C 224kPa 8:45  
136.2 C 224kPa 9:15  
136.2 C 224kPa 9:15  
VENTING 9:15  
AIR DRYING 10:16  
CYCLE COMPLETE 10:29  
-----

Digital Signature #  
8452E4F14116AB29  
-----