
Midmark M11 - v1.0.5
Total Cycles: 1141

Name: RITTER2
Sterilizer ID: V2313228
05/31/2022 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 83.7 F 0.0
4:00 78.0 F 0.0
6:00 98.1 F 0.1
8:00 137.4 F 2.9
10:00 163.6 F 6.6
12:00 187.8 F 10.4
14:00 225.2 F 7.3
16:00 228.1 F 7.3
18:00 236.8 F 11.2
20:00 231.7 F 7.0
22:00 242.4 F 12.6
24:00 239.7 F 10.8
26:00 251.5 F 16.9
28:00 262.2 F 23.0
30:00 272.2 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 29.6
2:00 272.7 F 30.0
3:00 271.5 F 29.5
4:00 272.8 F 29.7

Min 271.4 F 29.3
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 228.5 F 0.0
10:00 215.5 F 0.0
15:00 203.7 F 0.0
20:00 201.0 F 0.0
25:00 206.8 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

05/31/2022 08:00 AM
ACCEPTED: X
REJECTED:
Operator: U02