
Midmark M11 - v1.0.5
Total Cycles: 3187

Name: RITTER2
Sterilizer ID: V2313228
12/05/2023 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 161.0 F 2.2
4:00 183.2 F 6.2
6:00 203.8 F 10.8
8:00 222.9 F 5.4
10:00 226.8 F 10.5
12:00 231.7 F 7.7
14:00 242.7 F 13.5
16:00 244.1 F 12.6
18:00 256.4 F 19.3
20:00 267.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.3
1:00 272.6 F 29.9
2:00 271.4 F 29.4
3:00 272.4 F 29.7
4:00 271.4 F 29.6

Min 271.1 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 202.8 F 0.0
10:00 200.4 F 0.0
15:00 200.1 F 0.0
20:00 198.2 F 0.0
25:00 195.9 F 0.0
30:00 194.1 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:17
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

12/05/2023 05:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____