
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002257
15:44 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:14
STERILIZING 3:11
136.4 C 221kPa 3:11
136.1 C 224kPa 3:41
136.3 C 225kPa 4:11
136.4 C 223kPa 4:41
136.2 C 223kPa 5:11
136.4 C 222kPa 5:41
136.4 C 226kPa 6:11
136.4 C 226kPa 6:41
136.4 C 226kPa 6:41
VENTING 6:41
AIR DRYING 7:33
CYCLE COMPLETE 37:33

Digital Signature #
317E24951A9306BF
