

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2512

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2023 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
2:00 190.4 F 2.2  
4:00 188.2 F 6.3  
6:00 211.1 F 11.3  
8:00 223.7 F 6.3  
10:00 236.9 F 11.6  
12:00 234.4 F 8.9  
14:00 238.7 F 10.2  
16:00 246.2 F 13.8  
18:00 258.3 F 20.4  
20:00 269.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.7 F 29.7  
2:00 271.7 F 29.7  
3:00 272.2 F 29.5  
4:00 271.8 F 29.9

Min 271.2 F 29.2  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 219.6 F 0.0  
10:00 237.6 F 0.0  
15:00 232.2 F 0.0  
20:00 226.0 F 0.0  
25:00 224.9 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 21:52  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:32  
CYCLE COMPLETE

06/14/2023 09:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_