

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3863

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2024 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 80.9 F 0.0  
4:00 80.1 F 0.0  
6:00 115.3 F 1.5  
8:00 145.4 F 4.4  
10:00 171.4 F 7.7  
12:00 191.3 F 11.5  
14:00 208.9 F 3.5  
16:00 221.8 F 6.8  
18:00 229.6 F 10.4  
20:00 232.4 F 8.7  
22:00 238.6 F 11.0  
24:00 235.1 F 8.5  
26:00 249.1 F 15.0  
28:00 260.3 F 21.4  
30:00 270.4 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.8  
1:00 272.3 F 29.5  
2:00 272.7 F 29.1  
3:00 271.9 F 29.1  
4:00 273.0 F 29.5

Min 271.6 F 28.7  
Max 273.1 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
5:00 225.1 F 0.0  
10:00 190.9 F 0.0  
15:00 189.0 F 0.0  
20:00 185.0 F 0.0  
25:00 182.3 F 0.0  
30:00 180.8 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:45  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:08:51  
CYCLE COMPLETE

08/15/2024 09:18 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_