
Midmark M11 - v1.0.5
Total Cycles: 3279

Name: RITTER2
Sterilizer ID: V2313228
01/04/2024 03:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.5 F 0.0
2:00 157.9 F 2.2
4:00 172.2 F 5.8
6:00 195.1 F 10.2
8:00 214.1 F 6.5
10:00 225.4 F 9.2
12:00 233.7 F 10.2
14:00 241.1 F 12.2
16:00 240.8 F 11.3
18:00 253.9 F 17.9
20:00 265.6 F 24.9
22:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.7 F 29.8
2:00 271.8 F 29.7
3:00 272.6 F 29.8
4:00 271.6 F 29.5

Min 271.6 F 29.3
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 236.2 F 0.0
10:00 218.4 F 0.0
15:00 213.7 F 0.0
20:00 195.8 F 0.0
25:00 192.2 F 0.0
30:00 190.7 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

01/04/2024 04:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____