
Midmark M11 - v1.0.0
Total Cycles: 1063

Name: RITTER 1
Sterilizer ID: V2178198
07/28/2021 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.2 F 0.0
2:00 129.4 F 0.6
4:00 143.0 F 2.7
6:00 168.3 F 6.0
8:00 191.5 F 10.1
10:00 216.8 F 5.4
12:00 224.9 F 7.2
14:00 236.2 F 11.9
16:00 234.8 F 8.6
18:00 241.6 F 10.8
20:00 245.9 F 13.0
22:00 258.0 F 19.4
24:00 268.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 272.5 F 28.8
2:00 271.9 F 28.1
3:00 272.2 F 28.7
4:00 272.3 F 28.5

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 225.7 F 0.0
10:00 187.3 F 0.0
15:00 184.5 F 0.0
20:00 209.0 F 0.0
25:00 198.7 F 0.0
30:00 192.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

07/28/2021 09:21 AM
ACCEPTED: X
REJECTED:
Operator: U01