
Midmark M11 - v1.0.0
Total Cycles: 2430

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 07:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 80.3 F 0.0
4:00 84.2 F 0.0
6:00 111.5 F 1.2
8:00 141.5 F 3.8
10:00 175.0 F 6.8
12:00 186.3 F 10.4
14:00 210.0 F 6.1
16:00 215.6 F 6.0
18:00 226.1 F 9.8
20:00 238.0 F 10.5
22:00 236.9 F 10.4
24:00 235.1 F 8.2
26:00 247.1 F 14.0
28:00 258.3 F 20.0
30:00 268.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.1
2:00 271.6 F 28.7
3:00 272.3 F 29.1
4:00 271.5 F 28.7

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 215.4 F 0.0
10:00 235.4 F 0.0
15:00 228.3 F 0.0
20:00 223.5 F 0.0
25:00 223.2 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

10/14/2022 08:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____