
Midmark M11 - v1.0.0
Total Cycles: 4237

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2024 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.2 F 0.0
2:00 109.4 F 1.0
4:00 128.5 F 3.2
6:00 170.8 F 6.4
8:00 190.1 F 10.8
10:00 213.1 F 3.5
12:00 218.4 F 6.5
14:00 226.6 F 10.3
16:00 233.6 F 8.3
18:00 239.8 F 10.7
20:00 235.4 F 8.0
22:00 247.8 F 13.9
24:00 258.6 F 19.8
26:00 268.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.6 F 29.2
2:00 271.6 F 28.5
3:00 272.6 F 28.7
4:00 272.6 F 29.2

Min 271.0 F 28.0
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 223.2 F 0.0
10:00 199.7 F 0.0
15:00 220.3 F 0.0
20:00 222.5 F 0.0
25:00 230.5 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 27:05
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:05:41
CYCLE COMPLETE

01/26/2024 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____