

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001591  
14:31 11/17/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 4:41  
STERILIZING 6:10  
136.0 C 220kPa 6:10  
136.3 C 222kPa 6:40  
136.4 C 223kPa 7:10  
136.4 C 223kPa 7:40  
136.4 C 223kPa 8:10  
136.5 C 223kPa 8:40  
136.4 C 223kPa 9:10  
136.3 C 222kPa 9:40  
VENTING 9:40  
AIR DRYING 10:40  
CYCLE COMPLETE 40:40  
-----

Digital Signature #  
3AA3B2C1A8761689  
-----